

REPORT NUMBER

**18-275-4096**

REPORT DATE  
**Oct 02, 2018**

RECEIVED DATE  
**Sep 27, 2018**

SEND TO  
**11132**



13611 B Street • Omaha, Nebraska 68144-3693 • (402) 334-7770  
www.midwestlabs.com

**PAGE 1/2**

ISSUE DATE  
**Oct 02, 2018**

**SCHMIT GRETEMAN ASSOC INC  
KEN EFFERDING  
14927 INDUSTRIAL RD  
OMAHA NE 68144-**

**REPORT OF ANALYSIS**

For: (11132) SCHMIT GRETEMAN ASSOC INC  
BANNER  
A-B WATER STICKNEY, SD

Analysis	Level Found		Reporting			Analyst-Date	Verified-Date
	As Received	Units	Limit	Method			
Sample ID: <b>AB WATER</b> Lab Number: <b>2853083</b>							
Lead (total)	74.4	mg/kg	20.0	EPA 6010	ras7-2018/09/28	bab2-2018/10/02	

All results are reported on an AS RECEIVED basis., ppm = parts per million, ppm = mg/kg

For questions please contact:

Rob Ferris  
Account Manager  
rferris@midwestlabs.com (402)829-9871

The result(s) issued on this report only reflect the analysis of the sample(s) submitted.

Our reports and letters are for the exclusive and confidential use of our clients and may not be reproduced in whole or in part, nor may any reference be made to the work, the results, or the company in any advertising, news release, or other public announcements without obtaining our prior written authorization.

REPORT NUMBER

**18-275-4096**

REPORT DATE  
**Oct 02, 2018**

RECEIVED DATE  
**Sep 27, 2018**

SEND TO  
**11132**



13611 B Street • Omaha, Nebraska 68144-3693 • (402) 334-7770

www.midwestlabs.com

**PAGE 2/2**

ISSUE DATE  
**Oct 02, 2018**

**SCHMIT GRETEMAN ASSOC INC  
KEN EFFERDING  
14927 INDUSTRIAL RD  
OMAHA NE 68144-**

**REPORT OF ANALYSIS**

For: (11132) SCHMIT GRETEMAN ASSOC INC  
BANNER  
A-B WATER STICKNEY, SD

**Detailed Method Description(s)**

**ME 042**

Analysis follows MWL ME 042 which is based on EPA 6010b, Inductively Coupled Plasma (ICP). A light emission technique where prepared samples are injected into a high energy plasma that forces the elements in the injected sample to emit light energies which are proportional to the level of minerals and metals present. The light is then detected and correlated to the levels of minerals and metals in the original sample.

The result(s) issued on this report only reflect the analysis of the sample(s) submitted.

**Our reports and letters are for the exclusive and confidential use of our clients and may not be reproduced in whole or in part, nor may any reference be made to the work, the results, or the company in any advertising, news release, or other public announcements without obtaining our prior written authorization.**